

poz. 68 RT. 2016



INSTITUTUL DE PROIECTARE
INSTALATII PETROLIERE
STR. DILIGENTEI NR.18
PLOIESTI

SPECIFICATIE DE PROIECT

NR. 36-0467/R.03-0

DIVERSE APARATE

PAG. 1 DIN 1

COMANDA 157/2002 Ad 21/2003 Ob 04 PT/2 A4
BENEFICIAR S.C. PETROTEL LUKOIL S.A.
INSTALATIA REFORMARE CATALITICA

REV	DATA	INTOCMIT	VERIFICAT	APROBAT
0	02.04	DCA	DCR	BDA
1				
2				
3				
4				

ANALIZOARE DE UMIDITATE

SIMBOL: 04-AR-107

QUANTITY: 1

DESTINATIE: GAZ RECICLAT REFORMARE CATALITICA

PRESSURE: MAX.20 BAR

TEMPERATURE: MAX.50°C

COMPOZITIE

MOL. %

HCl	0,0001 TO 0,0015
H2	87,5466
CH4	4,50345
H2O	0,0005 TO 0,0025
C2H6	3,5130
C3H8	2,1851
C4H10	1,0769
C5H12	0,4113
C6+	0,7609

UTILITATI

PUTERE ELECTRICA:

220V, 50Hz 2 FAZE

AER INSTRUMENTAL:

3,94 BARG

NOTE: APARATELE CORESPUND SPECIFICATIILOR NR.75-0665/SU

CTSR

DRAGOS, CORMICA